

## Visit Report Sheet

<b>C1027</b>  AIG Control Systems Ltd Otter Pool Barn West Stowford Barton Barn, West Stowford, Ilfracombe Devon, England EX34 8PJ  07970 538998	<b>Visit Details</b>  Visit Type: Service Order No: 1479/ Date/Time: 04/03/2025 15:24 Time Completed: 15:39
<b>Job Details</b>  Service due,	<b>Lift Details</b>  Asset /Location: Ground Floor Living Room Make: Medi-Tec Model: Classic Serial: 2000172843054 S.W.L. 175kg Installed:
<b>Completion Notes</b>  Service of ground floor hoist.  <b>Parts Used</b>	<b>Further Work Required</b>

### VAT Exemption Certificate - Eligibility declaration by an individual

I declare that i am chronically sick or disabled and suffering from \_

I am receiving a qualifying Product or Service/Maintenance visit supplied to me for my domestic or personal use. I claim relief from Value Added Tax under group 12 of schedule 8 of the Value Added Tax Act 1994. The work has been completed.

**Customer Name:** AIG Control Systems Ltd

**Customer Sign:**

**Date:** 04/03/2025

**Rep Name:** Mark Pringle

**Rep Sign:**



**Date:** 04/03/2025



Service



## Service Checks

### Customer

AIG Control Systems Ltd  
Otter Pool Barn  
West Stowford Barton Barn,  
West Stowford,  
Ilfracombe  
Devon, England  
EX34 8PJ

**C1027**

### Visit Details

Visit Code: J1479  
Order No: TVMC/04/03/2025  
Consultant: Same As Site

Make: Medi-Tec  
Model: Classic  
Serial: 2000172843054

### Equipment Examined

Check Item	Status	Remarks
EIN number if present/Asset Number	N/A	
What is S.W.L. (Kg)?	Pass	175kg

### Examination Points


Check Item	Status	Remarks
Mast & leg fixings	Pass	
Ancillary fixings	Pass	
Up and down control	Pass	
Condition of lifting boom	Pass	
Actuator mounting bolts	Pass	
Life cycle of actuator (use calculator to approximate usage)	Pass	
Lift cycle number (approximate usage if unable to see)	Pass	781
Reset the service light (Molift)	N/A	
Retaining pins	Pass	
Spreader bar	Pass	
Traverse castors & brakes (measurement of castor if required)	Pass	
Lubricate where necessary	N/A	
Abnormal noise or resistance	Pass	
Mains wiring to charger	Pass	
Low voltage wiring	Pass	
Battery test 1 readings in v & Ah- Serial Number of Battery	Pass	12.87V 4.3Ah
Battery test 2 readings in v & Ah- Serial Number of Battery	Pass	12.91V 4.0Ah
Charger test readings v	Pass	27V DC
P.A.T. Test - Pass/Fail	N/A	
Up limit switch	Pass	
Down limit switch	Pass	
Emergency lowering manual	Pass	
Powered Emergency Lowering	N/A	

Emergency stop switch	Pass	
Hoist cover and head protection cover	N/A	
User instructions present	Pass	
Labelling inc SWL & contact numbers	Pass	
Weight test (kg)	N/A	
Pictures taken of weight test?	N/A	

### Consent Request

Check Item	Status	Remarks
Does client give consent for any pictures taken to be used on social media?	N/A	

**Rep Name:** Mark Pringle

**Rep Sign:** 

**Date:** 04/03/2025

