

Visit Report Sheet

<p style="text-align: right;">C1027</p> <p>AIG Control Systems Ltd Otter Pool Barn West Stowford Barton Barn, West Stowford, Ilfracombe Devon, England EX34 8PJ</p> <p>07970 538998</p>	<p style="text-align: center;">Visit Details</p> <p>Visit Type: Service Order No: 1363/ Date/Time: 04/03/2025 15:24 Time Completed: 15:39</p>
<p>Job Details</p> <p>Service due</p>	<p style="text-align: center;">Lift Details</p> <p>Asset /Location: 1st Floor Landing Make: Medi-Tec Model: Classic Serial: 2000171743054 S.W.L. 175kg Installed:</p>
<p>Completion Notes</p> <p>Service of 1st floor mobile hoist.</p> <p>Parts Used</p>	<p style="text-align: center;">Further Work Required</p>

VAT Exemption Certificate - Eligibility declaration by an individual

I declare that i am chronically sick or disabled and suffering from _

I am receiving a qualifying Product or Service/Maintenance visit supplied to me for my domestic or personal use. I claim relief from Value Added Tax under group 12 of schedule 8 of the Value Added Tax Act 1994. The work has been completed.

Customer Name: AIG Control Systems Ltd

Customer Sign:

Date: 04/03/2025

Rep Name: Mark Pringle

Rep Sign:



Date: 04/03/2025



Service



Service Checks

Customer

C1027

AIG Control Systems Ltd
Otter Pool Barn
West Stowford Barton Barn,
West Stowford,
Ilfracombe
Devon, England
EX34 8PJ

Visit Details

Visit Code: J1363
Order No: TVMC/04/03/2025
Consultant: Same As Site

Make: Medi-Tec
Model: Classic
Serial: 2000171743054

Equipment Examined

Check Item	Status	Remarks
EIN number if present/Asset Number	N/A	
What is S.W.L. (Kg)?	Pass	175kg

Examination Points


Check Item	Status	Remarks
Mast & leg fixings	Pass	
Ancillary fixings	Pass	
Up and down control	Pass	
Condition of lifting boom	Pass	
Actuator mounting bolts	Pass	
Life cycle of actuator (use calculator to approximate usage)	N/A	
Lift cycle number (approximate usage if unable to see)	Pass	687
Reset the service light (Molift)	N/A	
Retaining pins	Pass	
Spreader bar	Pass	
Traverse castors & brakes (measurement of castor if required)	Pass	
Lubricate where necessary	N/A	
Abnormal noise or resistance	Pass	
Mains wiring to charger	Pass	
Low voltage wiring	Pass	
Battery test 1 readings in v & Ah- Serial Number of Battery	Pass	12.62V 4.4Ah
Battery test 2 readings in v & Ah- Serial Number of Battery	Pass	12.67V 4.1Ah
Charger test readings v	Pass	27V DC
P.A.T. Test - Pass/Fail	N/A	
Up limit switch	Pass	
Down limit switch	Pass	
Emergency lowering manual	Pass	
Powered Emergency Lowering	N/A	

Emergency stop switch	Pass	
Hoist cover and head protection cover	N/A	
User instructions present	Pass	
Labelling inc SWL & contact numbers	Pass	
Weight test (kg)	N/A	
Pictures taken of weight test?	N/A	

Consent Request

Check Item	Status	Remarks
Does client give consent for any pictures taken to be used on social media?	N/A	

Rep Name: Mark Pringle

Rep Sign: 

Date: 04/03/2025



