

LOLER Report

<p style="text-align: right;">C1027</p> <p>AIG Control Systems Ltd Otter Pool Barn West Stowford Barton Barn, West Stowford, Ilfracombe Devon, England EX34 8PJ</p> <p>07970 538998</p>	<p style="text-align: center;">Visit Details</p> <p>Visit Type: LOLER Order No: 94539/ Date/Time: 14/11/2025 12:01 Time Completed: 12:48</p>
<p>Job Details</p> <p>LOLER inspection visit due, 2025</p>	<p style="text-align: center;">Lift Details</p> <p>Asset /Location: Ground Floor Bedroom Make: OpeMed Model: Elite 2 Compact Serial: 17160734 S.W.L. 200kg Date of Manufacture 01/01/2016 Installed: Date of Last Examination 21/06/2024 Next LOLER due</p>
<p>Completion Notes</p> <p>Serviced hoist. Batteries and charge rate are good. Batteries held charge after weight test. Lift tape has signs of wear and creasing which will need monitoring on next service. Cleaned all charge pickups including gates. Checked track and end stops. Weight tested to 190kg through gates and into bathroom. Reset service light. Replaced 1 missing white cap on spreader bar linkage. Inspection passed</p> <p>Parts Used</p>	<p style="text-align: center;">Further Work Required</p>

VAT Exemption Certificate - Eligibility declaration by an individual

I declare that i am chronically sick or disabled and suffering from _

I am receiving a qualifying Product or Service/Maintenance visit supplied to me for my domestic or personal use. I claim relief from Value Added Tax under group 12 of schedule 8 of the Value Added Tax Act 1994. The work has been completed.

Customer Name: AIG Control Systems Ltd

Customer Sign:

Date: 14/11/2025

Rep Name:Wayne Mav

Rep Sign: 

Date: 14/11/2025



Custom 16

LOLER Checks

<p>Customer</p> <p>AIG Control Systems Ltd Otter Pool Barn West Stowford Barton Barn, West Stowford, Ilfracombe Devon, England EX34 8PJ</p>	<p>C1027</p>	<p>Visit Details</p> <p>Visit Code: J94539 Order No: TVMC/14/11/2025/Elite 2 Compact Consultant: AIG Control Systems Ltd Contact: Tel:</p>
		<p>Make: OpeMed Model: Elite 2 Compact Serial: 17160734</p>

Equipment Examined

Question	Answer
Asset Number if present:	na
What is Manufacture Date?	2016
What is S.W.L. (Kg)?	200kg
Sling Inspection	n

Type of Examination

Question	Answer
Initial Examination	y
Examination after Reinstallation	n
Examination within a 6 -month Interval (9.3.a.i)	y
Examination In accordance with a written scheme	y
Exceptional Examination(Please Give Details)	weight tested to 190kg

Examination Points

Question	Answer
Battery Voltage and Amps (1)- Serial Number of Battery	12.60v 2.5ah
Battery Voltage and Amps (2)- Serial Number of Battery	12.63v 2.5ah
Check Charger & Docking Points	27.2v
Check Directional Controls/Function	y
Check Emergency/ Controlled Lowering	y
Check Labelling including SWL/ Service Dates	y
Are User Instructions Present?	y
Weight test/Load Amount (kg)	190kg
Pictures taken of weight test?	y
Limit/ Safety Switches	y
Track Position Related to Lifting Task	y
Track Fixings	y
Secondary Track	y
Track End Stops	y
Gate system Electrical/Mechanical	electrical
Hoist Rollers	y
Motor Carriage	y

Turntable	na
Safety Bar and Covers	y
Safety Bar Roll Pin Checked	y
Retaining Pins	y
Split Pins in Carrier Trolley (Robin CTH)	na
Track Joins/ Interlocks including Gates	y
Check lifting tape for signs of wear	slight wear and creasing
Abnormal Noise or Resistance	n
Dangerous Defects Found?	n
If Yes Please detail items which are safety critical now	n
Potentially dangerous defects (not safety critical) Found?	n
If Yes Please detail for monitoring	lift tape wear

Consent Request

Question	Answer
Does client give consent for pictures to be used on social media?	na

Rep Name: Wayne May

Rep Sign: 

Date: 14/11/2025





